



NAME S80653
 EXPNO 400
 PROCNO 1
 Date_ 20190816
 Time_ 14.53
 INSTRUM spect
 PROBHD 5 mm dual H1/C
 PULPROG zg30
 TD 65536
 SOLVENT DMSO
 NS 8
 DS 0
 SWH 8620.689 Hz
 FIDRES 0.131541 Hz
 AQ 3.8011379 sec
 RG 203
 DW 58.000 usec
 DE 6.50 usec
 TE 300.0 K
 D1 1.00000000 sec
 TD0 1

===== CHANNEL f1 =====
 NUC1 1H
 P1 13.22 usec
 PL1 -3.00 dB
 PL1W 20.99974442 W
 SFO1 400.1319006 MHz
 SI 32768
 SF 400.1300059 MHz
 WDW EM
 SSB 0
 LB 0.30 Hz
 GB 0
 PC 1.00