



NAME S63431
 EXPNO 400
 PROCNO 1
 Date_ 20191226
 Time_ 14.20
 INSTRUM spect
 PROBHD 5mm

PULPROG zg30
 TD 65536
 SOLVENT CDCl3
 NS 8
 DS 0
 SWH 8620.689 Hz
 FIDRES 0.131541 Hz
 AQ 3.8011379 s
 RG 203
 DW 58.000 μs
 DE 6.50 μs
 TE 300.0 K
 D1 1.00000000 s
 TD0 1

===== CHANNEL f1 =====
 NUC1 1H
 P1 13.22 μs
 PL1 -3.00 dB
 PL1W 20.99974442 V
 SFO1 400.1319006 MHz
 SI 32768
 SF 400.1300200 MHz
 WDW EM
 SSB 0
 LB 0.30 Hz
 GB 0
 PC 1.00