



上海源叶生物科技有限公司
Shanghai yuanye Bio-Technology Co., Ltd
电话: 400-666-5481 传真: 021-55068248
网址: www.shyuanye.com
邮箱: 3008007435@qq.com

质量检验报告

CERTIFICATE OF ANALYSIS

产 品 名 称：透析胎牛血清 Product Name:Diluted fetal bovine serum		抽 检 数：5 Sample Size		批 号：LM499151 Batch No.LM499151	
保质期：36个月					
参 照 标 准：CP2025 Specification					
生化检测 Biochemical Test	检查项目 Test Item	检验方法 Method	标准范围 Specification	分析结果 Result	
	无菌检验 Sterility Test	培养法 Culture Method	阴性 Negative	阴性 Negative	
	支原体 Mycoplasma	培养法 Culture Method	阴性 Negative	阴性 Negative	
	pH值 pH	酸碱度计法 pH Meter Method	6.8-8.0	7.5	
	血红蛋白 Hemoglobin	氰化高铁法 HiCN Method	≤0.05(W/V)	0.03	
	总蛋白 Total Protein	钨酸沉淀法 Tungstic Acid Method	3.5-5.0%(W/V)	4.5	
	内毒素 Endotoxin	凝胶法 Gel-clot Method	≤10EU/ml	4	
物理检测 Physical Test	颜 色 Color	观察外观 Visual Inspection	淡黄色 Light yellow	正 常 Normal	
	气味 Odor	闻气味 Olfactory Examination	特殊腥味 Characteristic odor	正常 Normal	
	性状 Appearance	观察外观 Visual Inspection	液体 Liquid	正 常 Normal	
结论 Conclusion	本产品符合CP2025标准 This product complies with CP2025.			合 格 / P a s s	
备注 Remark	1、本品经活性炭吸附后用进口微孔滤芯除菌过滤，在-20℃条件下保存。 This product is sterile-filtered through imported microporous filter cartridges after activated carbon adsorption, and stored at-20℃. 2、本产品最好采用逐步解冻法，解冻过程中需要随时轻摇，使温度和血清成份均匀 This product is best thawed gradually,shake gently throughout thawing to keep the temperature and serum components uniform.				

Quality Assurance: Xiaoyong Tan

Quality Control: Yunfei Zhu

质检专用章